RETURN TO SERVICE PROCEDURE TABLE OF CONTENTS
RETURN TO CD-ROM TABLE OF CONTENTS

Dräger medical
A Dräger and Siemens Company

Field Service Procedure

Euro Short Rail Assembly
Anesthesia

Revision -
10/28/03
SP00296

Field Service Procedure — Euro Short Rail Assembly

GCX Mounting Assembly Installation Guide

Purpose: The purpose of this document is to provide instruction for the mounting of the Euro Short Rail Assembly to Dräger Medical Anesthesia Machines.

Parts Reference: The following parts and hardware are included with this installation kit. (See illustrations for parts, hardware not shown):

Tools Required for This Procedure: None

Item #	Description	Qty
1	Euro Short Rail Assembly	1



SP296001

Fig. 1 GCX Euro Short Rail Assembly

1 INSTALLATION PROCEDURE: Mounting the Euro Short Rail Assembly (left side)

1. Insert the Slide (1) in the GCX channel (with thumb screw (2) at bottom) and move the Euro Short Rail Assembly to the desired height.
2. Using the thumb screw (2) to secure the Assembly in the channel.
3. Clamp accessory onto rail.

SP296.fm 28.10.03
All rights reserved. Copyright reserved.



Fig. 2 Euro Short Rail Assembly

All rights reserved. Copyright reserved.

Fig. 3 Euro Short Rail Assembly with Regulator Mounted (REF)

2 INSTALLATION PROCEDURE: Mounting the Euro Short Rail Assembly (right side)

1. Insert the Slide (1) in the GCX channel (with thumb screw (2) at top) and move the Euro Short Rail Assembly to the desired height.
2. Using the thumb screw (2) to secure the Assembly in the channel

SP296.fm 28.10.03
All rights reserved. Copyright reserved.



Fig. 4 Euro Short Rail Assembly

3. Clamp accessory onto rail.

All rights reserved. Copyright reserved.

Fig. 5 Euro Short Rail Assembly with Regulator Mounted (REF)

SP296.fm 28.10.03
All rights reserved. Copyright reserved.

Euro Short Rail Assembly — Field Service Procedure

All rights reserved. Copyright reserved.

In the United States:
Draeger Medical, Inc.
3136 Quarry Road
Telford, PA 18969
USA

Tel: (215) 721-5402
(800) 543-5047
Fax: (215) 721-5785

In all other countries:
Dräger Medical AG & Co. KGaA
Moislinger Allee 53 – 55
D-23542 Lübeck
Germany

Tel: (++49) (0) 1805-3723437
Fax: (++49) 451/882 - 3779

Euro Short Rail Assembly en
SP00296

Subject to change without notice.
Will not be replaced in the event of modifications.